

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Sathamangalam		<a href="#">Details About Robust Analysis</a>							
Lab No	10798		<a href="#">Detail About Monthly Summary</a>							
Month	March		<a href="#">Detail about SDI</a>							
Year	2023									
Constituent Group	Chemistry III									
<a href="#">Click on the analyte to view Graphical Data</a>			<a href="#">All Analyser Result</a>		<a href="#">Print</a>					
			Date of Result Entered : 20/03/2023							
			Date of Report Published : 04/04/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	115 mg/dL	3.45	0.31
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	139 mg/dL	4.84	0.28
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>