

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Shoologiri

Lab No 3641

Month August

Year 2023

Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SD	CV	L	U
						CV	SD					
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	170 mg/dL	-0.62	0.42		
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01		
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112 mg/dL	-0.39	0.22		

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Home	Modify My Profile	Submit Lab Results	View Lab Monthly Summary
Change Password	Home	Sign-out	

Ahne
MEDICAL OFFICER
 UPGRADED PRIMARY HEALTH CENTRE
 SHOOLAGIRI-635117.
 KRISHNAGIRI DISTRICT.