

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Royakottai
 Lab No: 12324
 Month: August
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 18/08/2023

Date of Report Published : 01/09/2023

Sl. No.	Analyte	Method / Principle Name	Result	Target	Delta	Delta Z	Delta X	Delta Y	Delta Z	Delta X	Delta Y
1	GLUCOSE III	GOD-POD III	1841	175.58	5.13	9.00	201 mg/dl	2.82	0.42		
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01		
3	CHOLESTROL III	CHOD-PAP III	1764	113.80	4.07	4.64	106 mg/dL	-1.68	0.22		

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

- [Monthly Address](#)
- [Monthly Methods](#)
- [Monthly Summary](#)
- [Change Password](#)

M. K. S.
MEDICAL OFFICER
 Govt. Primary Health Centre,
 RAYAKOTTAI - 635 116.