



## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO: SBS/CL/23/14997</b>		<b>MEDICAL DEVICES</b>	<b>Page No:1 of 1</b>	
Issue Date	13-10-2023			
SRF No & Date	SRF/23/00921-0001 & 17-10-2023			
Receipt Date	17-10-2023			
Calibration Date	17-10-2023			
Calibration Due	16-10-2024			
<b>Customer Name &amp; Address</b>				
GOVERNMENT PRIMARY HEALTH CENTRE, ERAGUDI-621001, TRICHY DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH	
Range	MULTI	Sr. No	ATCD1271220RBK	
Resolution	MULTI	Identification No	NA	
DUC Condition	SATISFACTORY	Location	LABORATORY	
<b>Environmental Conditions &amp; Calibration Procedure Details</b>				
Environmental Details	Temperature: 25.6 ° C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY

#### RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	99	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)
	<5000μA for B, BF, CF	203	7.4
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500μA for B, BF, CF	214	7.3


#### REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
(Calibration Engineer)  
K.SATHYAMOORTHY



  
 Technical Manager  
C.SHANMUGARAJ

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.