



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

|                   |                               |                            |            |
|-------------------|-------------------------------|----------------------------|------------|
| Lab Name          | Govt. PHC, Thorapalli         | Lab No                     | 12316      |
| Constituent Group | Chemistry III                 | Date of Result Entered :   | 20/06/2023 |
| PT item           | Lyophilized human serum based | Date of Report Published : | 04/07/2023 |

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |       | Your Value | SDI  | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|-------|------------|------|------|
|       |                |                                       |  |                    |        | CV           | SD    |            |      |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1823               | 286.63 | 5.17         | 14.82 | 297 mg/dL  | 0.70 | 0.69 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1848               | 2.57   | 6.64         | 0.17  | 2.6 mg/dL  | 0.18 | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1804               | 113.73 | 4.46         | 5.07  | 115 mg/dL  | 0.25 | 0.24 |

| SDI Range                       | Interpretation                           |
|---------------------------------|--|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

**LAB ADDRESS :**

Govt. PHC, Thorapalli  
 THE MEDICAL OFFICER, Hosur, Krishnagiri  
 Thorapalli  
 TAMILNADU635109

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*