

MONTHLY SUMMARY

Govt. PHC, Eppodumvendran

10633

April

2023

Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

All Analyser Result

Print

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
OSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	115 mg/dL	-0.62	0.34
TININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
ESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	126 mg/dL	2.49	0.28

SDI Range	Interpretation
00 to +1.00	Excellent.
.01 to ±2.00	Good.
.01 to ±2.99	Accept with caution. Warning Signal.
3.0	Unacceptable performance. Action Signal.

Home	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Primary	Change Password	Home	Sign-out