



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt UGPHC, Vadakadu Lab No 7155
Constituent Chemistry III Date of Result Entered : 20/09/2023
Group Lyophilized human serum based Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	302 mg/dL	-0.59	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	114 mg/dL	0.09	0.20
SDI Range										
				Within -1.00 to +1.00						
				Within ±1.01 to ±2.00						
				Within ±2.01 to ±3.99						
				Beyond ±3.0						
Interpretation										
				Excellent.						
				Good.						
				Accept with caution. Warning Signal.						
				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt UGPHC, Vadakadu
The Medical Officer, Vadakadu, Pudukkottai District
Vadakadu
TAMILNADU622304

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EOAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EOAS reports is confidential
CMC EOAS does not sub contract any components
***** End of Report *****