



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name: Govt. UGPHC, Vadakadu Lab No: 7155
 Constituent Group: Chemistry III Date of Result Entered: 20/04/2023
 PT Item: Lyophilized human serum based Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	Participants		Your Value	SDI	U
					CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21 mg/dL	-2.70	0.34
2	CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	1.9 mg/dL	-1.49	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79 mg/dL	0.24	0.28
SDI Range				Interpretation					
Within -1.00 to +1.00				Excellent.					
Within ±1.01 to ±2.00				Good.					
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.					
Beyond ±3.0				Unacceptable performance. Action Signal.					

LAB ADDRESS :
 Govt. UGPHC, Vadakadu
 The Medical Officer, Vadakadu, Pudukkottai District
 Vadakadu
 TAMILNADU622304

Paula Christudoss
Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in Dr. Pamela Christudoss
 Contact Number: 0416-2283102 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****