



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. UGPHC, Vadakadu Lab No: 7155
 Constituent: Chemistry III Date of Result Entered: 20/06/2023
 PT Item: Lyophilized human serum based Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	282 mg/dL	-0.31	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	96 mg/dL	-3.50	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. UGPHC, Vadakadu
 The Medical Officer, Vadakadu, Pudukottai District
 Vadakadu
 TAMILNADU622304

Coordinator Contact Details:
 Email: cinqc@cmcvellore.ac.in Dr. Pamela Christodoss
 Contact Number: 0416-2283102 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****