



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name	PATHOWYND CLINICAL LABORATORY	Lab No	0202
Constituent Group	HbA1c	Date of Result Entered :	14/09/2023
PT item	Lyophilized human whole blood based	Date of Report Published :	05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Biorad (D10 / D100 / Variant Turbo)	702	6.94	3.78	0.26	6.9 %	-0.15	0.02

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

PATHOWYND CLINICAL LABORATORY
ARUNAVA DAS 29-A, BALARAM GHOSH STREET
KOLKATA
WEST BENGAL 700004

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not subcontract any components

***** End of Report *****