



Bio-Rad
Laboratories (India) Private Limited
(A wholly owned subsidiary)

3rd, DLF IT Park I, Block – A,
8 Major Arterial Road, New-town,
Rajarhat, Kolkata – 700156
Tel. : (91)-33-22288564, 22268565
E-mail : sales.india@bio-rad.com
Tech Support : 1800 180 1224, 09873177477
www.bio-rad.com



ISO 9001:2015

www.tuv.com
ID 210527112

GSTIN : 19AAACB3202A1ZK
CIN No. : U32109DL1996PTC078494

DOCUMENT NO: SER-QR-09

Date: 08.05.2023

To,
Pathowynd Clinical Laboratory
Kolkata

Subject: Comprehensive Maintenance Contract of D10 HBA1C Analyzer

Dear Sir/Madam,

We would like to take this opportunity to thank you for using the Bio-Rad's **D10 HBA1C Analyzer System**. Please find enclosed quotation for Comprehensive Annual Maintenance Contract (CMC) including Spares with Services for **D10** installed at **Pathowynd Clinical Laboratory**, which has been tailored to suit your day to day operations. Listed below are some of services offered to ensure successful and efficient operation of your instruments.

- **Priority Service Response**
- **Telephone Service Support**
- **Scheduled Preventive Maintenance Calls**
- **Remote Diagnostics Support**

Kindly send one of the copies duly signed as a token of your approval along with payment in the form of NEFT/DD, favoring "Bio-Rad Laboratories (India) Pvt Ltd, Payable at Gurgaon. Please find the NEFT details as mentioned:

Beneficiary Name	Bio-Rad Laboratories (India) Pvt. Ltd.	Account No.(Current A/C)	5622420817
Bank Name	JP Morgan Chase Bank. New Delhi Branch	IFSC	CHASOINBX02 ('0' - denotes zero)
Branch	Units Nos – 401 & 402, 4 th Floor, Block – A, Hyatt Regency Delhi, Bhikaji Cama Place. New Delhi - 110066	9 Digit MICR Code	110500002

Invoice will be raised from below address:

Bio-Rad Laboratories (India) Private Limited
DLF IT Park I, Block – A, 3rd floor,
8 Major Arterial Road, New Town,
Kolkata – 700156

Thanking you and assuring you of our best services at all the times.

Sincerely yours

For Bio-Rad Laboratories (India) Pvt. Ltd .

Saumyabrata Ghosh
Sr. Account Manager

Note: You are requested to provide your GST Registration Certificate Copy & PAN card copy along with the Purchase Order copy.



Bio-Rad
Laboratories (India) Private Limited
(A wholly owned subsidiary)

3rd, DLF IT Park I, Block - A,
8 Major Arterial Road, New-town,
Rajarhat, Kolkata - 700156
Tel. : (91)-33-22268564, 22268565
E-mail : sales.india@bio-rad.com
Tech Support : 1800 180 1224, 09873177477
www.bio-rad.com



ISO 9001:2015

www.tuv.com
© 2006/07/11/13

GSTIN : 19AAACB3202A1ZK
CIN No. : U32109DL1996PTC078494

DOCUMENT NO: SER-QR-09

ANNUAL MAINTENANCE CONTRACT
(Comprehensive)

Name & Address of Customer: To, Pathowynd Clinical Laboratory Kolkata	Date: 8 th May, 2023 Agreement No.: BRI/CDG/K/CMC/PCL/05-2023 Validity: 60 Days Commencement Date : 20/05/2023 Termination Date : 19/05/2024 Renewal Date : 20/05/2024
---	--

Configuration/Tag #	Instruments	Location	Price-(INR)	Years	Total Price-(INR)
	D10 HBA1C ANALYSER (DC2L761106-DJ)	Pathowynd Clinical Laboratory	85000	1	85000
Total Amount					85000
(HSN Code: 9987) GST @ 18%					15300
One Lakh Three Hundred Only					100300

Signed on Behalf of Customer by	Signed on behalf of Bio-Rad Laboratories (India) Pvt. Ltd.
Name :	Name : Saumyabrata Ghosh
Title :	Title : Sr. Account Manager
Place :	Place : Kolkata
Date :	Date : 08-05-2023

Authorized Signatory

Saumyabrata Ghosh

- Note: 1. This Agreement is subjected to the terms and conditions overleaf (Service Plan BR/SER/CDG/19/TC & Standard).
2. Spare Parts & Services will be covered & this offer does not include third party items such as computer, printer & UPS etc.

BHARAT SCIENTIFIC HOUSE

144, B.B.Ganguly Street, 1st Floor, Room No.- 7,

Kolay Market, KOLKATA - 700012.

PH. : 03322276066 (M) : 9831034122

E-Mail : contactbsh1986@gmail.com

GSTIN : 19AALFB2845D1ZI State : 19 West Bengal PAN: AALFB2845D D.L.No: 12305SW 12112 SBW

TAX INVOICE

ORIGINAL For Recipient

PATHOWIND CLINICAL LABORATORY

29A, BALARAM GHOSH STREET

KOLKATA-700004

PH. : 25301329

(M) : 9830443540 Paym

GSTIN/ UIN :

STATE :

D.L.No :

Invoice No : BSH/23-24/JUN532

Invoice Dt : 14/06/2023

Credit days : 14/06/2023

Order No :

Order Dt :

Vendor Name : BHARAT SCIENTIFIC HOUSE

Vendor Code :

Service Period :

IRN:

Tax is Payable on Reverse Charge Basis (Yes/No) : No

ACK No.:

ACK Dt:

Sr No	HSN SAC	Product Description	Qty	Unit	Rate	CGST %	SGST %	Amount
1	998719	SERVICE BIO-RAD D10 CMC. SR.NO:DC2L761106 PERIOD:20.05.2023 TO 19.05.2024	1	PCS	85000.00	9	9	85000.00

1

85000.00

GST%	TAXABLE AMT	SGST	CGST	CGST	SGST	Amount
18.00%	of 85000.00=	7650.00+	7650.00	CGST	SGST	7650.00

TOTAL 85000.00 7650.00 7650.00

Round Off


Punjab National Bank, Paikpara Branch
A/c.No.0104250000725, IFS Code: PUNB0010420

NET AMOUNT

100300.00

[In Words] : One Lakh Three Hundred Only.

- (1) Any Discrepancy in quantity should be intimated at the time of delivery.
- (2) All interest will be charges @ 18% p.a. If Bill is not paid with from the date of presentation.
- (3) All disputes are Subject to Kolkata Jurisdiction only.



E. & O. E.
For BHARAT SCIENTIFIC HOUSE
[Signature]
Authorised Signatory



BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	PATHOWYND CLINICAL LABORATORY
Address:	29A, BALARAM GHOSH STREET, KOLKATA, WEST BENGAL
Province-Postal code	700004
Contact Name	Mr. ARUNAVA DAS

Calibration Certificate

Date <small>(dd mmmm, yyyy)</small>	11/10/2023	Instrument Model	D-10 Hb Analyzer	Instrument Serial Number	DC2L761106-DJ
---	------------	-------------------------	------------------	---------------------------------	---------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
VOLUMETRIC FLASK	N/A	N/A

Specified Range@ 1.5ml/Minute	Measured Value
1.425 – 1.575 ml/minute	1.48ml/min

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
FLUKE179	36030141	24/03/2023

Specified Range @ 35°C	Measured Value
33.4°C – 36.4°C	34.8

Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.019 AU

The next calibration is due on	11/10/2024	<small>(dd mmmm, yyyy)</small>
--------------------------------	------------	--------------------------------

For Bio-Rad Laboratories India Pvt Limited
Authorized Signatory:

SM Choudhury