




CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. Block PHC, Pudhuchaatiram Lab No 7023
 Constituent Group Chemistry III Date of Result Entered : 18/07/2023
 PT Item Lyophilized human serum based Date of Report Published : 07/08/2023

| Sl.No | Analyte | Method / Principle Name | Analyser Name | No of Participants | DV | Participants | | Your Value | SDI | LI |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1876 | 109.35 | 5.55 | 6.07 | 108 mg/dL | -0.22 | 0.28 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1850 | 5.83 | 6.00 | 0.35 | 5.9 mg/dL | 0.20 | 0.02 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1776 | 113.00 | 4.12 | 4.66 | 114 mg/dL | 0.21 | 0.22 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. Block PHC, Pudhuchaatiram

"The Block Medical Officer Govt.Primary Health Centre, Puduchathiram Cuddalore District-608501"

Pudhuchaatiram

TAMILNADU608501

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****