

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Puthur.T	Details About Robust Analysis								
Lab No	7242	Detail About Monthly Summary								
Month	July	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data						All Analyser Result		Print		
Date of Result Entered : 18/07/2023										
Date of Report Published : 07/08/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	105 mg/dL	-0.72	0.28
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	184	51.41	11.14	5.73	48 mg/dL	-0.60	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6 mg/dL	0.49	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out