

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Puthur.T	<a href="#">Details About Robust Analysis</a>								
Lab No	7242	<a href="#">Detail About Monthly Summary</a>								
Month	June	<a href="#">Detail about SDI</a>								
Year	2023									
Constituent Group	Chemistry III									
<a href="#">Click on the analyte to view Graphical Data</a>						<a href="#">All Analyser Result</a>		<a href="#">Print</a>		
Date of Result Entered : 20/06/2023										
Date of Report Published : 04/07/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	280 mg/dL	-0.45	0.69
2	<a href="#">UREA III</a>	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	192	32.69	12.96	4.24	27 mg/dL	-1.34	0.61
3	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
4	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	111 mg/dL	-0.54	0.24
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>