

ALLWIN MEDICAL SYSTEMS

No. 2/4, 2nd Street, 1st Floor, Jayalakshmi Nagar, Kattupakkam, Chennai - 600 056, Tamilnadu, India,

Cell: 9443663366

Email: allwinmedicalsystems@gmail.com

CALIBRATION CERTIFICATE

Date: 29.09.2023

AWMS43C23 Certificate No Calibration Name & Address

Government Primary Health Centre

Mullendiram

Details of Device Under Calibration (DUC)

Description

Least Count

Range

: Centrifuge

DUC Condition

: 0-3000 RPM

: ±10 RPM

Make: Labotech Model: BDi152

Sr.No: 20230419 Location: Lab

Environmental Conditions & Calibration Procedure Details

Relative humidity 49%RH Temperature: 25°C **Environmental Details** Calibration Done at **ON SITE** 06.03.23 Sample Calibration date

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS	DEVIATION	EXPANDED UNCERTAINTY(±) 25 RPM
		RPM		
1	1200	1192	-8	Ok
2	1800	1796	-4	Ok
3	2400	2400	Ok	Ok
4	3000	2996	-4	Ok

Remarks

- 1. This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory
- 2. The user Should be determine the suitability of the instrument for its intended use
- 3. The Recalibration interval should be determine on the User requirement.
- 4. The results Stated in this Certificate relate only to the item Calibrated
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00

6. Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By