

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Thiruchurpettai

Lab No

12932

Constituent Group

Chemistry III

Date of Result Entered :

04/03/2023

PT item

Lyophilized human serum based

Date of Report Published : 04/04/2023

CI No	Analyta	Method / Principle	Analyzor Namo	No of	DV	Participants		Your	SDI	U
SI.No	Analyte	Name .	Analyzer Name	Participants	υv	CV	SD	Value	וחפ	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	90.72 mg/dL	-0.11	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.41 mg/dL	0.66	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	113 mg/dL	0.36	0.28

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Thiruchurpettai Thiruchurpettai phc Thiruvannamalai district Thiruchurpettai TAMILNADU606803

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Thiruchurpettai Lab No 12932

Constituent Group Chemistry III Date of Result Entered : 05/04/2023

PT item Lyophilized human serum based Date of Report Published : 06/05/2023

CLNo	SING Method / Princip		Analyzar Nama No of		DV	Participants		Your	SDI	
SI.No	Analyte	Name .	Analyzer Name	Participants	DV	CV	SD	Value	וחפ	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	133.2 mg/dL	1.91	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.57 mg/dL	2.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	123 mg/dL	1.97	0.28

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Thiruchurpettai Thiruchurpettai phc Thiruvannamalai district Thiruchurpettai TAMILNADU606803

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Thiruchurpettai Lab No 12932

Constituent Group Chemistry III Date of Result Entered : 23/05/2023

PT item Lyophilized human serum based Date of Report Published : 05/06/2023

CLNG	SING Analysis Method / F			Partici DV		Participants		CDI	U	
SI.No	Analyte	Name .	Analyzer Name	Participants	DV	CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	135 mg/dL	-4.89	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.5 mg/dL	-1.97	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	105 mg/dL	-1.65	0.27

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Thiruchurpettai Thiruchurpettai phc Thiruvannamalai district Thiruchurpettai TAMILNADU606803

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Thiruchurpettai Lab No 12932

Constituent Group Chemistry III Date of Result Entered: 03/06/2023

PT item Lyophilized human serum based Date of Report Published : 04/07/2023

CLNo	SLNa Method / P			No of DV		Participants		Your	SDI	U
SI.No	Analyte	Name .	Analyzer Name	Participants	DV	CV	SD	Value	וחפ	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	247.7 mg/dL	-2.63	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.44 mg/dL	-0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	92.11 mg/dL	-4.26	0.24

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Thiruchurpettai Thiruchurpettai phc Thiruvannamalai district Thiruchurpettai TAMILNADU606803

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Thiruchurpettai Lab No 12932

Constituent Group Chemistry III Date of Result Entered : 20/07/2023

PT item Lyophilized human serum based Date of Report Published : 07/08/2023

CLNo	SI.No Analyte Method / Pri		le Analyzor Namo No of	DV Partic		Participants		SDI	U	
SI.NO	Analyte	Name	Analyzer Name	Participants	DV	CV	SD	Value	ועפ	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	109.7 mg/dL	0.06	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.42 mg/dL	-1.17	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112.6 mg/dL	-0.09	0.22

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Thiruchurpettai Thiruchurpettai phc Thiruvannamalai district Thiruchurpettai TAMILNADU606803

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name Govt. PHC, Thiruchurpettai Lab No 12932

Constituent Group Chemistry III Date of Result Entered: 13/08/2023

PT item Lyophilized human serum based Date of Report Published : 01/09/2023

CLNo	Analyta	Method / Principle		No of	DV	Participants		Your	SDI	U
SI.No	Analyte	Name .	Analyzer Name	Participants	DV	CV	SD	Value	וחפ	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	168.1 mg/dL	-0.83	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.55 mg/dL	-1.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	110.3 mg/dL	-0.76	0.22

SDI Range	Interpretation					
Within -1.00 to +1.00	Excellent.					
Within ±1.01 to ±2.00	Good.					
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.					
Beyond ±3.0	Unacceptable performance. Action Signal.					

LAB ADDRESS:

Govt. PHC, Thiruchurpettai Thiruchurpettai phc Thiruvannamalai district Thiruchurpettai TAMILNADU606803

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss