

VIEW LAB MONTHLY SUMMARY

Lab Name **ASAVLEE ,DR APARNA'S PATHOLOGY LABORATORY**[Details About Robust Analysis](#)Lab No **6878**[Detail About Monthly Summary](#)Month **September**[Detail about SDI](#)Year **2023**Constituent Group **Chemistry II**[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)Date of Result Entered : **07/09/2023**Date of Report Published : **05/10/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	CPC (Turbochem 100 / 240 Prime)	98	337.58	11.92	40.24	305 mg/dL	-0.81	8.13
2	UREA II	Urease UV / GLDH II	CPC (Turbochem 100 / 240 Prime)	79	38.00	10.79	4.10	35.4 mg/dL	-0.63	0.92
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	813	1.18	14.43	0.17	1.1 mg/dL	-0.47	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	CPC (Turbochem 100 / 240 Prime)	91	5.33	12.13	0.65	7 mg/dL	2.59	0.14
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1299	5.64	10.51	0.59	5.3 g/dL	-0.57	0.03
6	ALBUMIN II	BCG - Colorimetric II	CPC (Turbochem 100 / 240 Prime)	86	3.32	8.16	0.27	3 g/dL	-1.18	0.06
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	CPC (Turbochem 100 / 240 Prime)	84	9.79	10.02	0.98	7.1 mg/dL	-2.74	0.21
8	CHOLESTROL II	CHOD-PAP II	CPC (Turbochem 100 / 240 Prime)	84	116.32	10.65	12.39	70 mg/dL	-3.74	2.70
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	CPC (Turbochem 100 / 240 Prime)	89	207.95	9.53	19.81	265 mg/dL	2.88	4.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.

Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out