



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - FEBRUARY 2023



Lab Name Govt.UGPHC, Jamnamaruthur
Constituent Group Chemistry III
PT Item Lyophilized human serum based

Lab No 7524
Date of Result Entered : 21/02/2023
Date of Report Published : 08/03/2023

SL No	Analyte	Manufacturer/ Principle Name	Analyser Name	No. of Participants	CV	XRD Limits		Test Value	SDI	V
						CV	TD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	172 mg/dL	1.66	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3 mg/dL	-1.31	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1788	112.79	5.23	5.90	109 mg/dL	-0.64	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution, Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt.UGPHC, Jamnamaruthur
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT.
Jamnamaruthur
TAMILNADU635703

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential

Neyyath
9/3/23
MEDICAL OFFICER,
Govt. Primary Health Cen
Jamunamarathur-63570
Javadhu Hills, T.V. Maia