



# G MULTI SERVICES


78, Yashob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 462.  
gmultiservice@gobi@gmail.com. Call : 96597 20654.  
Bio Medical Instruments Services

Er. A. Ibathul Rasid DECE., BE (ICE)  
Bio Medical Engineer

## Calibration Protocol & Report for Medical Equipment

Equipment : PIPETTE  
MAKE : FINNPIPETTE  
MODEL : 100-1000 $\mu$ L

PROTECT CLASS / TYPE : Risk Class II  
S. No. : BA/PI/002  
Location : GPHC, BARUR

S.No	Measurement Parameters		Status	Remarks
1	Electrical Test	(a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	----	-
2	Physical Condition Test	(a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	Performance Test	(e) Mean 99.2 $\mu$ l (f) A -0.5 $\mu$ l (g) Acc % -0.71 (h) CV% -0.36	OK OK OK Ok	
4	Calibration Test	min volume 100 $\mu$ l max Volume 1000 $\mu$ l (ELECTRONIC BALANCEING TEST) (e) A+or - 1.5, 5 $\mu$ l (f) S 1.5 $\mu$ l max Vol. 1000 $\mu$ l (g) Acc+ or - 1.50, 0.50 (h) CV% 0.50, 0.10	Ok Ok Ok Ok	Tested
Formulated By: Edition Research Laboratory		Issued By:	User Department:	
Endorsed By : Research team head				
Reviewed On: 26.09.23 Next Review : 25.09.24				

