

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Nammiyampattu	Details About Robust Analysis	
Lab No	7515	Detail About Monthly Summary	
Month	April	Detail about SDI	
Year	2023		
Constituent Group	Chemistry III		

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	132.3 mg/dL	1.78	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	1.9 mg/dL	-1.49	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	118.5 mg/dL	1.19	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார இலையர்
 இழுமாமல்கோள்-635703, த.ம.ந.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
----------------	----------------	--------------------	--------------------------