



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt.UGPHC, Jamnamaruthur Lab No 7524
 Constituent Group Chemistry III Date of Result Entered : 20/09/2023
 PT item Lyophilized human serum based Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	314 mg/dL	0.21	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	109 mg/dL	-1.07	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt.UGPHC, Jamnamaruthur

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THIRUVANNAMALAI DISTRICT.

Jamnamaruthur

TAMILNADU635703

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub-contract any components

***** End of Report *****

[Signature]
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 ஜமுனாமரத்தூர்-635703. தி.நா.