## SARAVANA MEDICO

Shop No. 41/50, Good Shed Road, 4th Street, Adambakkam, Chennai - 600 088 E-mail : saravanamedico2019@gmail.com, Mobile : 7358077155

				,		Page No	: 01 of 02		
CALIBRATION CERTIFICATE					Temp		: 25±4 °C		
					Humidity	: 30-75 RH%			
CERTIFICATE NO: LDM/CAL/UP		/CAL/UPHC	C/0323/SA01 ULR.NO:			Date of Issue	: 30.09.2022		
NAME AND ADDRESS OF THE CUSTOMER			Date of Receipt of item		: 25.09.2022				
The Medical Officer				Date of Calibration 3		30.092022	30.092022		
Government UPHC,				Suggested date of Calibration		: 29.09.2023			
Periyar Nagar ,Avadi				Ref. Documents		: As per Manufacturer specification			
				Cal. Procedure Ref. No.		: LDM-CP-MD19			
				Customer Ref No.		: Verbal			
			DES	SCRIPTION OF DUC	il 9			, j =	
Name of the DUC		: Semi Auto Analyser		Receiving Condition of the DUC		: Good			
Make		:		Location		: Laboratory			
Model		:		Calibrated at		: Onsite			
Sl.no.		:		Range		: Multi			
ID No.		:		Resolution		: Multi			
			DETAI	LS OF MASTER USED				,	
Standard Instrument		nt Used	Make / Model	ID No./ Sl.No.	Traceat	eable to / Certificate No.		Validity	
Electrical Safety Ana		nalyzer	Datrend /vPad Rugged 2	VRR20070452		2780622-E01		29-Sep-23	
The Maste	er are tracea	ble to Natio	onal / International stan						
VISUAL INSPECTION							Remarks		
SL.No.		Description					Pass	Fail	
1	Device Clea	Device Clean and Decontaminated					V		
2	Physical damage						V		
3 Power Cord check							V		

