

# SARAVANA MEDICO

Shop No. 41/50, Good Shed Road, 4th Street, Adambakkam, Chennai - 600 088  
E-mail : saravanamedico2019@gmail.com, Mobile : 7358077155

Certificate No.: E131221-1-1

Name & Address of the Customer

M/s. Upgraded Primary Health Center,  
Periyar Nagar Avadi

SRF No. : E131221-1  
Received on : 25.09.2022  
Date of Calibration : 30.09.2022  
Recom. Due Date : 29.11.2022

## Details of Test Instrument

|                         |                          |                      |                        |
|-------------------------|--------------------------|----------------------|------------------------|
| Description             | : Micropipette           | Range                | : 1 ml                 |
| Make                    | : ---                    | Operating Range      | : ---                  |
| Model No. / Type        | : ---                    | Resolution           | : ---                  |
| Serial No.              | : ---                    | Accuracy             | : As per specification |
| ID No. / Asset No.      | : --                     | Location             | : ---                  |
| Calibrated at           | : Lab                    | Condition On Receipt | : Satisfactory         |
| Work Instruction No.    | : EMTPL/WI/MECH-14       | Based on Standard    | : ISO 8655-6 : 2002    |
| Environmental Condition | : Temperature: 23±1.5 °C | Relative Humidity    | : 50±10 %RH            |

## Details of Master Instrument

| Description                   | ID No.       | Traceability          | Valid till |
|-------------------------------|--------------|-----------------------|------------|
| Electronic Semi Micro Balance | EMTPL-EWB-01 | 2022-23/EQN/210252-01 | 29.11.2022 |

The Masters are traceable to National/International Standard

## CALIBRATION DATA MECHANICAL CALIBRATION (Volume)

| Test Reading | Standard Reading | Deviation | Expanded Uncertainty (±) |
|--------------|------------------|-----------|--------------------------|
| ml           | ml               | ml        | ml                       |
| 1            | 0.996            | -0.004    | ---                      |

## Remarks :

1. This certificate refers to the values for the above instrument at the time of calibration and under the above stated conditions.
2. The Calibration Certificate shall not be reproduced except in full without the written approval of EMTPL
3. The Reported Expanded Uncertainty is at 95.45% Confidence level., coverage factor k=2

Calibrated by :  
Quality Manager



Authorised by :  
Technical Manager

