



2 & 3, Devi Illam, Priya Nagar,
Telungupalayam Pirivu,
Coimbatore -641 039, TamilNadu.

T. ARUN PRASAD, M.Tech. MBA. (Ph.D)
Managing Director
Consultant Biomedical Engineer

GSTIN: 33APLPA 1828A1ZM

Calibration Protocol & Report for Medical Equipment

Equipment: **MICRO PIPETTE**
Make : FINNPIPETTE
Model: 10-100 µL

Protect Class/Type: B
S.No : RW10206
Location : PERIYAMUTHUR-GPHC

S.no	Measurement Parameters	Status	Remarks
I	Electrical test (a). oral method digital apparatus	Ok	Well
II	Physical Condition Test (a). Cabinet & Visible damages, if any (b). on/off switch, digital display, glass (c). Risk of mechanical damages, if any	No OK No	Checked ok
III	Performance Test (a), Glass visual clarity (b). oral sensor edge (c). Abnormalities (d). Measurement checkup (e). on/off functioning (f). deviations, if any	Normal Ok No Ok Well No	Spares replacement/ Checked & measured ok
IV	Calibration Test (a). digital Zero calibration: 1. zero degree (with freezer) 2. normal body temp (warmer/incubator) 3. 100 micro litre (water) (b). Measurement correlation test	Mercury zero level confirmed / Checked ok with std machine	Ensured zero level Fine
Formulated By: Einstein Research Laboratory		Issued By:	User Department:
Endorsed By: Research team head			
Reviewed On : 13.09.2023 Next Review: 12.03.2024			

Email: tcabsmedical@gmail.com Mobile: 9597089432, 9500741818