

CALIBRATION CERTIFICATE

Date : 03.10.2023

Calibration Name & Address		Certificate No	AWMS82C23
Government Primary Health Centre Ulundhai.			
Details of Device Under Calibration (DUC)			
Description	: Centrifuge	Make	: Remi
Range	: 0-3000 RPM	Model	: c 852
Least Count	: ± 10 RPM	Sr.No	: 16877
DUC Condition	:	Location	: Lab
Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25°C	Relative humidity	54%RH
Sample Calibration date	06.03.23	Calibration Done at	ON SITE

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS	DEVIATION	EXPANDED UNCERTAINTY(\pm) 25 RPM
RPM				
1	1200	1200	0	Ok
2	1800	1798	-02	Ok
3	2400	2400	0	Ok
4	3000	3000	0	Ok

Remarks

- 1.This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory
- 2.The user Should be determine the suitability of the instrument for its intended use
- 3.The Recalibration interval should be determine on the User requirement.
- 4.The results Stated in this Certificate relate only to the item Calibrated
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00
- 6.Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By



Authorized Signatory

