



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/1114

Page No: 1 of 1

Customer Name & Address *

GOVERNMENT PRIMARY HEALTH CENTRE, PANNANDUR-635123, KRISHI NAGIRI DISTRICT.	SRF No.	SRF/23/008/0-0007
	SRF Date	11-10-2023
	Date of Receipt	10-10-2023
	Date of Calibration	11-10-2023
	Due Date for Calibration	10-10-2024
	Issue Date	12-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	NA
Range	100µl	Model	NA
Resolution	FIXED	Material	PVC
Serial Number	NA	Operating Range	100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	21.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.4	Cal Procedure No	SBS/CP/ML/08

Standard Used

SL No.	Description	ID No. / SL No.	Cert/Scale No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP/22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl										
SL No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (z)
		1	100	99.89	99.88	99.89				

Remarks

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade) as specified in ISO 3696.

Calibrated By:

M. P. Lakshmi
(Calibration Engineer)
MAZHUBAN



Authorised by:

L. Siva
(Quality Manager/Chief Executive)
CSIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

Office: 1st Floor, Poomagal Main Road, Ekattivihangal, Chennai - 600 032. Tel: 044 - 2225 2017
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com

