



2 & 3, Devi Illam, Priya Nagar,  
Telungupalayam Pirivu,  
Coimbatore -641 039, TamilNadu.

**T. ARUN PRASAD, M.Tech. MBA. (Ph.D)**  
Managing Director  
Consultant Biomedical Engineer

GSTIN: 33APLPA 1828A1ZM

**Calibration Protocol & Report for Medical Equipment**

Equipment: **CENTRIFUGE**  
Make: REMI  
Model: REMI BDI-152

Protect Class/Type: Risk class 1  
S.No :7015  
Location : Alapattii-UGPHC

S.no	Measurement Parameters	Status	Remarks
I	<b>Electrical test</b> (a). Safety Standards (range: 210-240 V) (b). Equipotential ground & Earthing (c). Risk of shock/fire Hazards, if any (d). Chassis & Insulation Resistance (e). Earth & Chassis Leakage Current (f). Battery condition & charging status	230 volts 0 to 5 volts no zero ohms NA	Stabiliser/ups Wih in ranges
II	<b>Physical Condition Test</b> (a). Cabinet & Visible damages, if any (b). Spo2 probe, clip, extn cable, Power cord & accessories (c). Risk of mechanical damages, if any)	No ok No	Checked ok
III	<b>Performance Test</b> (a). On/off switch (b). Speed regulator switch & functional tests (c). Running test (d). Abnormalities (e). Display checkup (f). Keypad/indicators & connectors	Normal Ok accurate ok Ok Well	Perfect Perfect Ok
IV	<b>Calibration Test</b> a). Vibration test b). Speed regulation gradual hike/ c). Running test with load d). Display checkup f). timer test	Done ok	Perfect
<b>Formulated By:</b> Einstein Research Laboratory		<b>Issued By:</b>	<b>User Department:</b>
<b>Endorsed By:</b> Research team head			
Reviewed On :11/10/2023 Next Review:10/04/2024			

Email: [tcabsmedical@gmail.com](mailto:tcabsmedical@gmail.com) Mobile: 9597089432, 9500741818\