



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Kuppanasaripatti	<a href="#">Details About Robust Analysis</a>
Lab No	12705	<a href="#">Detail About Monthly Summary</a>
Month	April	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	110 mg/dL	-1.31	0.34
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	169	42.32	10.68	4.52	43 mg/dL	0.15	0.70
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.2 mg/dL	0.23	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	38	4.58	9.59	0.44	4.6 mg/dL	0.05	0.14
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	108 mg/dL	-0.62	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Kuppanasaripatti	<a href="#">Details About Robust Analysis</a>
Lab No	12705	<a href="#">Detail About Monthly Summary</a>
Month	August	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	181 mg/dL	0.60	0.42
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	174	28.29	15.20	4.30	22 mg/dL	-1.46	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	40	5.43	10.18	0.55	5.5 mg/dL	0.13	0.17
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	111 mg/dL	-0.60	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

**Lab Name** Govt. PHC, Kuppanasaripatti [Details About Robust Analysis](#)  
**Lab No** 12705 [Detail About Monthly Summary](#)  
**Month** July [Detail about SDI](#)  
**Year** 2023  
**Constituent Group** Chemistry III

[Click on the analyte to view Graphical Data](#)
[All Analyser Result](#)
[Print](#)

**Date of Result Entered :** 20/07/2023  
**Date of Report Published :** 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	120 mg/dL	1.76	0.28
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	184	51.41	11.14	5.73	47 mg/dL	-0.77	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.9 mg/dL	0.20	0.02
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	30	5.35	9.89	0.53	5.9 mg/dL	1.04	0.19
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out





### VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kuppanasaripatti

[Details About Robust Analysis](#)

Lab No 12705

[Detail About Monthly Summary](#)

Month June

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	282 mg/dL	-0.31	0.69
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	192	32.69	12.96	4.24	31 mg/dL	-0.40	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.3 mg/dL	-1.58	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	34	6.25	4.51	0.28	6.2 mg/dL	-0.18	0.10
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	118 mg/dL	0.84	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



Site Maintained by

**Swiftware Solutions**

Best Viewed at pixels 1024 X 768



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Kuppanasaripatti	<a href="#">Details About Robust Analysis</a>
Lab No	12705	<a href="#">Detail About Monthly Summary</a>
Month	May	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	179 mg/dL	-0.20	0.44
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	163	27.91	13.39	3.74	24 mg/dL	-1.05	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.6 mg/dL	-1.21	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	40	4.98	10.94	0.55	4.6 mg/dL	-0.70	0.17
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	116 mg/dL	0.24	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Kuppanasarihatti	<a href="#">Details About Robust Analysis</a>
Lab No	12705	<a href="#">Detail About Monthly Summary</a>
Month	September	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	310 mg/dL	-0.06	0.69
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	187	35.19	12.73	4.48	28 mg/dL	-1.61	0.66
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	35	6.72	4.36	0.29	7 mg/dL	0.96	0.10
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	115 mg/dL	0.32	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out