



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. PHC, B.Mutlur Lab No: 7635
Constituent Group: Chemistry III Date of Result Entered: 19/05/2023
PT item: Lyophilized human serum based Date of Report Published: 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	182.11 mg/dL	0.13	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.77 mg/dL	0.08	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	110.22 mg/dL	-0.75	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, B.Mutlur
The Medical Officer, Govt.Primary Health Centre, B.Mutlur, Cuddalore District
B.Mutlur
TAMILNADU608501

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Pamela Christudoss
மருத்துவ அலுவலர்
அ.ஆரம்ப சுகாதார நிலையம்
- 4.06.2023