



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name	Govt. PHC, Pavithram	Lab No	12428
Constituent Group	Chemistry III	Date of Result Entered :	19/08/2023
PT item	Lyophilized human serum based	Date of Report Published :	01/09/2023

Sl. No.	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Mean Value	SDI	CV	ST
						C.V	ST				
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	182 mg/dL	0.71	0.42	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.00	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	114 mg/dL	0.04	0.00	0.00

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Pavithram
CLINICAL LABORATORY
Pavithram
TAMILNADU 637021

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

[Handwritten Signature]
19/08/23
MEDICAL
Govt. Primary
Pavithram
Namakkal