



CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE AMATHUR, VIRUDHUNAGAR, TAMIL NADU-626001		CERTIFICATE NO	TICS/AMA/23-24/023
		EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE MODEL	THERMO SCIENTIFIC 10-100 µL
CALIBRATION DATE	13/07/2023	SERIAL NO	Rw01367
CALIBRATION DUE DATE	12/07/2024	INSTALLED DEPARTMENT	LAB

- 1. Have more dust in machine outside Yes No
- 2. Any Physical damaged in machine Yes No
- 3. Machine inside Cleaning (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSP22/07/884-01	27-JUL-2023

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.4	± 1	PASS
		99	99.3	± 1	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols. using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE
Calibrated By: *Per* Authorized Sign With Seal

