



## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO:</b> SBS/CL/23/14625	<b>MEDICAL DEVICES</b>	<b>Page No:</b> 1 of 1
Issue Date	14-10-2023	
SRF No & Date	SRF/23/00885-0001 & 13-10-2023	
Receipt Date	13-10-2023	
Calibration Date	13-10-2023	
Calibration Due	12-10-2024	

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
MANGALAKUDI-623308, RAMANATHAPURAM DISTRICT.

**Details of Device Under Calibration (DUC)**

Description : SEMI AUTO ANALYZER	Make & Model : ALPHA & ALPHA CHEM
Range : MULTI	Sr. No : BCAA17062608
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

**Environmental Conditions & Calibration Procedure Details**

Environmental Details	Temperature: 25.4°C	Relative Humidity	54% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

**ELECTRICAL SAFETY**

**RESULTS**

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 100	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µA for B, BF, CF	Measured values in µA 156	Uncertainty in % (±) 7.8
3	Enclosure Leakage <500µA for B, BF, CF	Measured values in µA 215	Uncertainty in % (±) 7.3


**REMARKS**

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
(Calibration Engineer)  
P.PRASANNA



  
Technical Manager  
C.SHANMUGARAJ

Chief Executive

### SUNSHINE BIOMEDICAL SOLUTIONS

No.68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/14626

Page No: 1 of 1

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE, MANGALAKUDI-623308, RAMANATHAPURAM DISTRICT.	SRF No.	SRF/23/00885-0002
	SRF Date	13-10-2023
	Date of Receipt	12-10-2023
	Date of Calibration	13-10-2023
	Due Date for Calibration	12-10-2024
	Issue Date	14-10-2023

**Details of Unit Under Calibration**

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW13645	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023


Z Factor: 1.00319

Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	100	99.94	99.96	99.97	99.95	99.90	99.96	-0.04	0.02	0.47	
		99.96	99.95	99.97	99.98	99.98					
2	500	499.95	499.96	499.95	499.97	499.98	499.96	-0.04	0.01	0.47	
		499.96	499.97	499.95	499.95	499.96					
3	1000	999.95	999.96	999.97	999.96	999.98	999.96	-0.04	0.01	0.47	
		999.95	999.96	999.97	999.97	999.97					

**Remarks**

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,   
 (Calibration Engineer)  
 M.RAGUL



Authorised by:  
  
 (Quality Manager/Chief Executive)  
 C.SIVABALAN

### SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/14627

Page No: 1 of 1

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE, MANGALAKUDI-623308, RAMANATHAPURAM DISTRICT.	SRF No.	SRF/23/00885-0003
	SRF Date	13-10-2023
	Date of Receipt	12-10-2023
	Date of Calibration	13-10-2023
	Due Date for Calibration	12-10-2024
	Issue Date	14-10-2023

**Details of Unit Under Calibration**

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW10276	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.7	Humidity (%RH)	52	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023


**Z Factor: 1.00319**

Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10.0	9.90	9.92	9.88	9.91	9.92	9.91	-0.09	0.01	0.47
		9.91	9.92	9.93	9.92	9.90				
2	50.0	49.78	49.76	49.74	49.72	49.74	49.74	-0.26	0.02	0.47
		49.76	49.75	49.73	49.72	49.71				
3	100.0	98.92	98.94	98.94	98.96	98.92	98.93	-1.07	0.02	0.47
		98.91	98.93	98.94	98.92	98.91				

**Remarks**

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade 2 as specified in ISO 3696.

Calibrated By,   
 (Calibration Engineer)  
 M.RAGUL



Authorised by:  
  
 (Quality Manager/Chief Executive)  
 C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.