

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. UPHC, Arcot Lab No 9367

Constituent Group Chemistry III Date of Result Entered : 22/01/2021

PT item Lyophilized human serum based Date of Report Published : 05/02/2021

CL No.	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	CDI	
SI.No						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1621	110.65	6.87	7.60	113 mg/dL	0.31	0.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1552	1.32	11.10	0.15	1.6 mg/dL	1.90	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1562	110.60	7.12	7.88	121 mg/dL	1.32	0.40

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. UPHC, Arcot

THE MEDICAL OFFICER, ARNI ROAD, TOPUKANNA, ARCOT

Arcot

TAMILNADU632503

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudoss
Dr. Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

********* End of Report *******