



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023

VELLORE CMC

FINAL REPORT

MAY - 2023



Lab Name

Govt. UPHC, Veerapandi

Lab No

10869

Constituent Group

Chemistry III

Date of Result Entered :

19/05/2023

PT item

Lyophilized human serum based

Date of Report Published :

05/06/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | CV | SD | Your Value | SDI | U |
|-----------------------|----------------|---------------------------------------|--|--|--------|------|------|------------|-------|------|
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1834 | 180.88 | 5.19 | 9.39 | 170 mg/dL | -1.16 | 0.44 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1854 | 1.76 | 7.48 | 0.13 | 1.7 mg/dL | -0.45 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1798 | 114.60 | 5.08 | 5.83 | 108 mg/dL | -1.13 | 0.27 |
| Interpretation | | | | | | | | | | |
| SDI Range | | | | | | | | | | |
| Within -1.00 to +1.00 | | | | Excellent. | | | | | | |
| Within ±1.01 to ±2.00 | | | | Good. | | | | | | |
| Within ±2.01 to ±2.99 | | | | Accept with caution. Warning Signal. | | | | | | |
| Beyond ±3.0 | | | | Unacceptable performance. Action Signal. | | | | | | |

MEDICAL OFFICER
UPHC: Veerapandi

Handwritten signature

Handwritten signature: Pamela Christudoss

LAB ADDRESS :
Govt. UPHC, Veerapandi
MEDICAL OFFICER, VEERAPANDI, PALAVANJIPALAYAM
Tirupur
TAMILNADU641608

Coordinator Contact Details:
Email: cfmqcc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential