




CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name: Govt. UPHC, Veerapandi Lab No: 10869
 Constituent Group: Chemistry III Date of Result Entered: 20/09/2023
 PT Item: Lyophilized human serum based Date of Report Published: 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	312 mg/dL	0.08	0.69
2	CREATININE III	Jaffes Kinetic - Alkalline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.4 mg/dL	1.05	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	116 mg/dL	0.55	0.20
SDI Range										
Within -1.00 to +1.00 Excellent.										
Within ±1.01 to ±2.00 Good.										
Within ±2.01 to ±2.99 Accept with caution. Warning Signal.										
Beyond ±3.0 Unacceptable performance. Action Signal.										
Interpretation										

LAB ADDRESS :
 Govt. UPHC, Veerapandi
 MEDICAL OFFICER, VEERAPANDI, PALAVANJIPALAYAM
 Tirupur
 TAMILNADU641608

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

P. Veerapandi
MEDICAL OFFICER
 UPHC, Veerapandi.