



MECOM SERVICES

Reg. Office
15/582B, ST MARY'S BUILDING, AM ROAD, SOUTH VAZHAKULAM, ALUVA,
COCHIN - 683105
Ph: 8586315926 / 25298 E Mail: mecomcalibration@gmail.com Web: www.mecomservices.in



REPORT ON

ELECTRICAL SAFETY TESTING/ PERFORMANCE ANALYSIS/ CALIBRATION

Report No: MS/TR/22/M14	Calibration Date: 10-08-2022
Page 1 Of 2	Calibration Due: 10-08-2023
1.1 CUSTOMER DETAILS	
Name and address of the organisation	PRIMARY HEALTH CENTER MUDALUR, SANTHANKULAM TOOTHUKKUDI TAMILNADU-628702
Reference and Date	AS PER THE MAIL CONFIRMATION
Test performed by	MECOM team-1

1.2 DESCRIPTION OF DEVICE UNDER TEST (DUT)

Nomenclature	MICROSCOPE
A. Manufactured by	LABOMED
B. Model	VISION-2000
C. Serial No.	CM/L9435282

1.3 CONDITION OF THE ITEM WHEN RECEIVED

No visible damage and in working order

1.4 Applicable Specification	IEC Specification IEC 60601-1
------------------------------	-------------------------------

1.5 Test Done	Electrical Safety Test
---------------	------------------------

1.6 STATUS

Manufactures Specification	Users Specification	Within Specification	Out of Specification	Calibration	Electrical Safety Test	Performance Analysis
✓	-	✓	-	-	✓	-

1.7 TRACEABILITY DETAILS OF INSTRUMENTS USED FOR TESTING

SL No	Name of the Instrument	Make	Model	Serial No	Cal Due
1.	ELECTRICAL SAFETY ANALYZER	FLUKE	ESA 612	3005050	MAY 2023



MECOM SERVICES

REGD. OFFICE
15/587B AT KIRARU BUILDING, AMERICAL, SATHANKULAM AVINAI,
CANCOR, ERUDUR
IN. ARKONNAM, 628702. TEL: 9442122222, 9442122222. FAX: 9442122222, 9442122222



REPORT ON
ELECTRICAL SAFETY TESTING/ PERFORMANCE ANALYSIS/ CALIBRATION

Report No: MS/TR/22/M13	Calibration Date 10-08-2022
Page 1 Of 2	Calibration Due 10-08-2023
1.1 CUSTOMER DETAILS	
Name and address of the organisation	PRIMARY HEALTH CENTER MUDALUR, SATHANKULAM TOOTHUKKUDI TAMILNADU-628702
Reference and Date	AS PER THE MAIL CONFIRMATION
Test performed by	MECOM team-1

1.2 DESCRIPTION OF DEVICE UNDER TEST (DUT)

Nomenclature	CENTRIFUGE
A. Manufactured by	BOI
B. Model	BOI
C. Serial No.	BC1112

1.3 CONDITION OF THE ITEM WHEN RECEIVED
No visible damage and in working order

1.4 Applicable Specification IEC Specification IEC 60601-1

1.5 Test Done Performance Test and Electrical Safety Test

1.6 STATUS

Manufactures Specification	Users Specification	Within Specification	Out of Specification	Calibration	Electrical Safety Test	Performance Analysis
✓		✓		✓	✓	✓

1.7 TRACEABILITY DETAILS OF INSTRUMENTS USED FOR TESTING

Sl No	Name of the Instrument	Make	Model	Serial No	Cal Due
1.	ELECTRICAL SAFETY ANALYZER	FLUKE	ESA 612	3005050	MAY 2023
2.	DIGITAL TACHOMETER	LUTRON	DT-2230	Q655879	JUN 2023