

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Kayalpattinam	Details About Robust Analysis	
Lab No	13000	Detail About Monthly Summary	
Month	August	Detail about SDI	
Year	2023		
Constituent Group	Chemistry III		
Click on the analyte to view Graphical Data			
		All Analyser Result	Print
Date of Result Entered : 19/08/2023			
Date of Report Published : 01/09/2023			

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	23	173.17	4.83	8.36	176 mg/dL	0.34	3.49
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	59	28.92	13.99	4.05	27 mg/dL	-0.47	1.05
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	21	1.74	6.34	0.11	1.71 mg/dL	-0.27	0.05
4	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	17	111.42	3.18	3.54	113 mg/dL	0.45	1.72

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out