

Customer Name: The Medical Officer
Address: Govt Primary Health Centre
Vellalur

Description of Device Under

Equipment Name	Micro pipette
Value	10ul-100ul
Serial No	YE224AX0183885

Calibration Standards Used :

Master Instrument	Model	Serial No	Date of Calibration	Calibration Due on
Electrical Compact Scale	YE224X	YE224AX0183885	26-09-2023	25-09-2024

Number of Mesurement	Channel 1 (Volume)		
	10ul	50ul	100ul
SRNO			
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.1
4	10.09	50.08	100.08
5	10.01	49.94	100.09
6	10.08	50.07	100.12

k.k.
SIGNATURE