

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Ponbethi

[Details About Robust Analysis](#)

Lab No 7150

[Detail About Monthly Summary](#)

Month July

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	115 mg/dL	0.93	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out