



Trust & Care



heartbeat

Medical Systems

CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 23.05.2023
		Calibration Valid: 22.05.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, URBAN PRIMARY HEALTH CENTRE, CORPRATION HOSPITAL, NANJUNDAPURAM, COIMBATORE.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	REMI
ii	Model	----
iii	Serial No	ZEAN22878
iv	Location	LABORTAORY

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM	Combined uncertainty in RPM in % rdg
1	1000	1000	2	1.12
2	1200	1201	2	1.23
3	1500	1502	4	1.56
4	2000	1999	1	1.175

Acceptable deviation is ± 5 RPM it depends on Max Speed Range



No. 23, 2nd Floor, jayasimmapuram,
Pappanaikenpalayam, Coimbatore – 37.

99400 28355, 97512 06106

88833 15662

hbmedicalsystem@gmail.com

hbmssales@gmail.com