



SUMMARY

Govt. PHC, Nakkambadi  
12398  
July  
2023  
Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/07/2023  
Date of Report Published : 07/08/2023

	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	134 mg/dL	4.06	0.28
II	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	107 mg/dL	-1.29	0.22

Range	Interpretation
> +1.00	Excellent.
0 ±2.00	Good.
0 ±2.99	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>



LAB SUMMARY

Govt. PHC, Nakkambadi  
12398  
August  
2023  
Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/08/2023  
Date of Report Published : 01/09/2023

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI
					CV	SD		
III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	167 mg/dL	-0.95
IV	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.65 mg/dL	-0.92
IV	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	114.2 mg/dL	0.09

Range	Interpretation
0 to +1.00	Excellent.
1 to +2.00	Good.
1 to +2.99	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>





MONTHLY SUMMARY

Govt. PHC, Nakkambadi  
12398  
June  
2023  
Department Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023  
Date of Report Published : 04/07/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
OSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	288 mg/dL	0.09	0.69
TININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.4 mg/dL	-0.99	0.01
ESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	110.2 mg/dL	-0.70	0.24

SDI Range	Interpretation
1.00 to +1.00	Excellent.
1.01 to ±2.00	Good.
2.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

Home	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Logout	Change Password	Home	Sign-out



Govt. PHC, Nakkambadi  
12398  
May  
2023  
Chemistry III

[Details About Robust Anal](#)  
[Detail About Monthly Sum](#)  
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

Date of Result Entered : 20

Date of Report Published :

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants	
					CV	SD
COSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39
TININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13
ESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83

SDI Range	Interpretation
1.00 to +1.00	Excellent.
±1.01 to ±2.00	Good.
±2.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

ess	Modify Methods	Submit Lab Results	View L
nary	Change Password	Home	Sign-o

Site Maintained by  
**Swiftware Solutions**  
Best Vie

/ SUMMARY

Govt. PHC, Nakkambadi [Details About Robust Analysis](#)

12398 [Detail About Monthly Summary](#)

April [Detail about SDI](#)

2023

up Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

ite	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value
					CV	SD	
I	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	124 mg/dL
E III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.6 mg/dL
DL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	117 mg/dL

Range	Interpretation
to +1.00	Excellent.
to ±2.00	Good.
to ±2.99	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*Site Maintained by*  
**vifware Solutions**

Best Viewed at p



LY SUMMARY

Govt. PHC, Nakkambadi  
12398  
September  
2023  
Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/09/2023  
Date of Report Published : 05/10/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	
					CV	SD			
Urea III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	291 mg/dL	-1.33	0
Urea III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0
Urea III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	112 mg/dL	-0.38	0

SI Range	Interpretation
0 to +1.00	Excellent.
1 to ±2.00	Good.
1 to ±2.99	Accept with caution. Warning Signal.
2	Unacceptable performance. Action Signal.

<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

