



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12728

Page No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, VEMBI-605203,VILLUPURAM DISTRICT.	SRF No. SRF Date Date of Receipt Date of Calibration Due Date for Calibration Issue Date	SRF/23/00590-0001 03-10-2023 03-10-2023 03-10-2023 02-10-2024 04-10-2023
--	---	---

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW09041	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.2	Humidity (%RH)	53	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10.0	9.87	9.86	9.88	9.88	9.88	9.87	-0.13	0.01	0.47
		9.88	9.88	9.87	9.86	9.87				
2	50.0	49.74	49.76	49.78	49.74	49.76	49.76	-0.24	0.02	0.47
		49.78	49.74	49.78	49.78	49.78				
3	100.0	99.84	99.86	99.87	99.87	99.88	99.87	-0.13	0.01	0.47
		99.87	99.88	99.86	99.87	99.88				


Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,


 (Calibration Engineer)
 M.ADHIBAN

Authorized by:


 (Quality Manager/Chief Executive)
 C.SIVABALAN



SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.