



CALIBRATION CERTIFICATE

No: SBS/CL/13/13909

Page No. 1 of 1

Name & Address

KENT PRIMARY HEALTH CENTRE,
TAPURAM-6(120), MAYILADUTHURAI DISTRICT.

SRF No.	SRF/23/000636-000
SRF Date	09-10-2023
Date of Receipt	07-10-2023
Date of Calibration	09-10-2023
Due Date for Calibration	08-10-2024
Issue Date	09-10-2023

Unit Under Calibration

No.	Micro Pipette	Make	ERBA BIOMET
	10-100µl	Model	NA
	1ml	Material	PVC
Serial No.	7066718	Operating Range	10-100µl
	NA	Condition of UDC	Good
	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temp (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbars)	1006	Water Temperature (°C)	23.8	Cal Procedure No.	SBS/CP/ML/08

Jewel

Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability
Electronic Weighing Balance	15112918	TVCSP122/12/2115-01	A&D & GH-232	National Standards

Z Factor: 1.0

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty U
		9.90	9.92	9.98	9.91	9.92				
1	10	9.91	9.92	9.93	9.92	9.90	9.91	-0.09	0.03	0.47
		49.78	49.76	49.74	49.72	49.74	49.74	-0.26	0.02	0.47
2	50	49.76	49.75	49.73	49.72	49.71	49.74	-0.26	0.02	0.47
		98.92	98.94	98.94	98.96	98.92	98.93	-1.07	0.02	0.47
3	100	98.91	98.93	98.94	98.92	98.91	98.93	-1.07	0.02	0.47

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.

4. The results stated in this certificate relate only to the item calibrated.

5. Equipment used for Calibration were calibrated & traceable to National & International Standards.

6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.

7. Calibration Liquid Used: Distilled or Deionized water conforming to Grade A as specified in ISO 3696.

Calibrated By,

M. Achikirup
(Calibration Engineer)

Authorised by:

A. S. Venkateswaran
(Quality Management Executive)