

**VIEW LAB MONTHLY SUMMARY**

<b>Lab Name</b>	Govt. UPHC, Periyar nagar	<b>Details About Robust Analysis</b>
<b>Lab No</b>	10902	<b>Detail About Monthly Summary</b>
<b>Month</b>	May	<b>Detail about SDI</b>
<b>Year</b>	2023	
<b>Constituent Group</b>	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 22/05/2023

Date of Report Published : 05/06/2023

Sl No	Analyte	Method Name	Principle Name	Analyser Name	No of Participants	CV	CV	CV	Value	SDI	U
1	GLUCOSE III	GOD-POD III		Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	198 mg/dL	1.82	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III		Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III		Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	104 mg/dL	-1.82	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Home</a>	<a href="#">About Us</a>	<a href="#">Services</a>	<a href="#">Contact Us</a>
<a href="#">Lab Reports</a>	<a href="#">Lab Results</a>	<a href="#">Lab Summary</a>	<a href="#">Lab Details</a>

*J. Mahalingam*