

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Periyar nagar
 Lab No: 10902
 Month: June
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

Sl No	Analyte	Method	Instrument	Count	Mean	SD	CV	SDI	Warning	Action
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	316.73 mg/dL	2.03	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.39 mg/dL	-1.05	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	101 mg/dL	-2.51	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good
Within +2.01 to +2.99	Accept with caution. Warning Signal
Beyond +3.0	Unacceptable performance. Action Signal.

Links

P. Subudhi