

PHC, Kadalaiyur

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

istry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 15/07/2023

Date of Report Published : 07/08/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
				CV	SD			
POD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	107.3 mg/dL	-0.34	0.28
affes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.33 mg/dL	1.43	0.02
POD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112.5 mg/dL	-0.11	0.22

Interpretation
Excellent.
Good.
Accept with caution. Warning Signal.
Unacceptable performance. Action Signal.

D. Jay
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
KADALAIYUR

