

PHC, Kadalaiyur

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

istry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 15/06/2023

Date of Report Published : 04/07/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
				CV	SD			
OD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	314.6 mg/dL	1.89	0.69
affes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.47 mg/dL	-0.58	0.01
HOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	114.8 mg/dL	0.21	0.24

Interpretation

Excellent.

Good.

Accept with caution. Warning Signal.

Unacceptable performance. Action Signal.

D. Gay
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 KADALAIYUR

Modify Methods

Submit Lab Results

View Lab Monthly Summary

Change Password

Home

Sign-out

