CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023

PC-1024

t. PHC, Kadalaiyur

nistry III

hilized human serum based

Lab No 12995

Date of Result Entered : 10/05/2023

Date of Report Published : 05/06/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
D-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	191.3 mg/dL	1.11	0.44
fes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.64 mg/dL	-0.91	0.01
OD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	113.9 mg/dL	-0.12	0.27

	Interpretation
	Excellent.
	Good.
	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
KADALAIYUR

dinator Contact Details:
il:clinqc@cmcvellore.ac.in
act Number: 0416-2283102

Scanned with OKEN Scanner