

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Regunathapuram  
 Lab No: 10886  
 Month: September  
 Year: 2023  
 Constituent Group: Chemistry III

Details About Robot Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result Print

Date of Result Entered: 20/09/2023  
 Date of Report Published: 03/10/2023

Sl.No	Analyte	Method / Principle name	Analyser Name	No of Pa Samples	CV	Pa Samples		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	329 mg/dL	1.21	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Phtrate III	Robonik (Semi automatic / Fully Automatic)	1832	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTRDL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	114 mg/dL	0.09	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution. Warning Signal
Beyond ±3.0	Unacceptable performance. Action Signal

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

*(Handwritten Signature)*  
**Asst. Surgeon**  
**Govt. Primary Health Centre**  
**Regunathapuram.**



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt PHC, Regunathapuram	<a href="#">Details About Robust Analysis</a>
Lab No	10886	<a href="#">Detail About Monthly Summary</a>
Month	August	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 26/08/2023

Date of Report Published : 01/09/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SO			
1	GLUCOSE III	GDH-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	0.00	165 mg/dL	-1.18	0.42
2	CREATININE III	Jaffee Kinetic - Alkaline Phosphate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.79	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	117 mg/dL	8.69	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Home	Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out	

**Asst. Surgeon**  
 Govt. Primary Health C.  
 Regunathapuram



Summary

Govt. PHC, Regunathapuram

10586

July

2023

Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result Print

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
				CV	SD			
GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.15	5.55	6.07	110 mg/dL	0.11	0.28
Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6 mg/dL	0.49	0.02
CHCO-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22

	Interpretation
+1.00	Excellent.
+2.00	Good.
+2.99	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

Modify Methods	Submit Lab Results	View Lab Monthly Summary
Change Password	Home	Sign-out

Maintained by Software Solutions

Best Viewed at pixels 1024 X 768

*(Signature)*  
**Asst. Surgeon**  
 Govt. Primary Health Co.  
 Regunathapuram

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Regunathapuram  
Lab No: 10566  
Month: June  
Year: 2023  
Constituent Group: Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Date of Result Entered: 21/06/2023


Date of Report Published: 04/07/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GCO-POD III	Robonik ( Semi automatic / Fully Automatic)	1923	286.63	5.17	14.92	291 mg/dL	0.29	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.9 mg/dL	1.93	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	112 mg/dL	-0.34	0.24


SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Modify Address	Change Password	Home	Sign-out
Yearly Summary			

Site Maintained by

 **Swiftware Solutions**

Best Viewed at pixels 1024 x 768

  
**Asst. Surgeon**  
Govt. Primary Health Centre  
Regunathapuram

Lab Name: Govt. PHC, Begunahapuram  
 Lab No: 0200  
 Month: April  
 Year: 2023  
 Consultant Group: Chemistry III

[Details About Robonik Analyser](#)  
[Detail About Monthly Summary](#)  
[Detail about SD](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 10/04/2023

Date of Report Published : 06/05/2023

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SD	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.46	6.04	7.21	114 mg/dL	0.76	0.30
2	CHLATERINE III	Jaffes Kinetic Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	6.09	0.18	2.4 mg/dL	1.37	0.01
3	CHOLESTEROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	112 mg/dL	0.07	0.20

SDR Range	Interpretation
Within -1.00 to +1.00	Excellent
Within +1.01 to +2.00	Good
Within +2.01 to +2.99	Accept with caution. Warning Signal
Beyond +3.0	Unacceptable performance. Action Signal.