

Date: 10-01-2023
Effective Date: 10-01-2023

Certificate of Calibration

Customer Name: Asavalec, Dr Aparna pathology Laboratory

Model : Automated Hematology Analyzer Sysmex KX-21

Serial No. : B1652

Calibration Done Date: 10.1.23

Next Calibration Due Date On or Before: 09-01-2024

Lab In-charge: . DR.Aparna Jairam

This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, HCT and PLT) according to the standard procedures provided by Sysmex Corporation, Japan.

The reference instruments used for value-assignment are managed by the traceability system in Sysmex Corporation and these are traceable to the International Standards, such as ICSH.

Calibration at site performed by
Engineer Name Vinit Vartak
Designation RSM
Transasia Bio-Medicals Ltd
Location Mumbai

Encl:

1. Certificate of Inspection
2. Assay Sheet of Calibrator SCS-1000
3. Printouts
4. Traceability & Uncertainty document